



Tryptic Soya Broth(SCDB)-Supplemented with 0.05% SPS (20mL, 50mL, 70mL)

Tryptic Soya Broth(SCDB) with 0.05% SPS is generally recommended for the detection of most aerobic /anaerobic bacteria and other fastidious microorganisms in blood.

Product Presentation:

Cat No.	Product description	Pack Size
22010030020	Tryptic Soya Broth(SCDB)-Supplemented with 0.05% SPS	20 mL
22010030050	Tryptic Soya Broth(SCDB)-Supplemented with 0.05% SPS	50 mL
22010030070	Tryptic Soya Broth(SCDB)-Supplemented with 0.05% SPS	70 mL

Principle and interpretation

Tryptic Soya Broth(SCDB) with 0.05% SPS use the indirect procedure for obtaining the blood specimen from the patient. Blood is withdrawn from the patient using a sterile needle and syringe and is added to the Blood Culture Bottle containing 0.05% Sodium Polyanethol Sulfonate (SPS). Sodium Polyanethol Sulfonate (SPS) is a polyanionic anticoagulant, which inhibits complement and lysozyme activity, interferes with phagocytosis and inactivates aminoglycosides. It is generally considered to enhance the rate and speed of bacterial isolations by counter-acting the bacterial inhibitors of human blood. After thorough and gentle mixing, the blood specimen is added generally in the ratio of 1 mL of blood to 10 mL of broth to the blood culture bottle. The bottle is incubated and observed for turbidity, colour change, hemolysis, gas formation or other evidence of microbial growth.

Composition

Ingredients

	Grams / Litre
Tryptone	15.00
Soya Peptone	5.00
Dextrose	2.50
Sodium Chloride	5.00
Di-Potassium Phosphate	2.50
SPS	0.50

Final pH (at 25°C) 7.3±0.2

**Formula adjusted, standardized to suit performance parameters

Type of specimen

Clinical samples: Blood

FACTORY & OFFICE

Plot No. D 76 , Five Star MIDC Area, Kagal.
Dist. Kolhapur -416216 (M.S.)India.
Email : oxalispvtltd@outlook.com

Telefax : 0231-2305072
Phone : 0231-2305062
Mobile : +91 8805867810



Specimen Collection and Preparation

- ✓ No special preparation of the patient is required prior to sample collection by approved techniques.
- ✓ The specimen must be collected using sterile techniques to reduce the chance of contamination.
- ✓ Collect 1:10 ratio blood to broth, For Adult approximately 5 mL of patient's blood when using the 50 mL, approximately 7 mL of patient's blood when using the 70 mL Blood culture medium bottle and approximately 1-3 mL when using the 20 mL Blood culture medium bottle and approximately 0.5-1.5 mL when using the 5-10 mL Blood Culture Medium bottle.
- ✓ Samples should be stored at 2°C-8°C if not tested immediately.
- ✓ Avoid using hemolysed samples for testing.

Directions

- ✓ Label the ready to use blood culture bottle.
- ✓ Remove the Aluminium foil cap. Disinfect the part of the rubber stopper which is now exposed.
- ✓ Draw patient's blood with the sterile or disposable needle and syringe as explained in specimen collection and disposable column.
- ✓ Transfer the blood sample immediately into the culture bottle by puncturing the rubber stopper with the needle and injecting the blood. Venting: Use sterile venting needle. Keep the bottle in an upright position preferably in a biological safety cabinet, place an alcohol swab over the rubber stopper and insert the venting needle with filter through it. Insertion and withdrawal of the needle should be done in a straight line.
- ✓ Discard the needle and mix the contents by gently inverting the bottle 2-3 times. Do Not vent the bottle for anaerobic cultures.
- ✓ Incubate at 35±2°C for 18-24 hours and further for seven days.

Storage and Stability

- ✓ Store the blood culture media in cool, dry place at 15°C-25°C away from direct light.
- ✓ Stability of the blood culture media is as per the expiry date mentioned on the label.

Quality Control

Appearance: Light amber coloured, clear solution without any precipitate.

Appearance on Addition of Blood: Cherry red coloured opaque solution without any blood clot.

Cultural Response: Cultural response is observed after an incubation of 18-24 hours at 35°C-37°C and subculturing is carried out on Blood Agar.

Organism (ATCC)	Growth	Subculturing on Blood Agar
<i>Staphylococcus aureus</i> (25923)	Good	Beta Haemolysis
<i>Streptococcus pyogenes</i> (19615)	Good	Beta Haemolysis
<i>Bacteroides vulgatus</i> (8482)	Good	-

Interpretation of Results

- ✓ Growth in the broth medium is indicated by the presence of turbidity. The bottles should be held for 7 days before reporting a negative blood culture.

Warranty

- ✓ This product is designed to perform as described on the label and package insert. The manufacturer disclaims any implied warranty of use and sale for any other purpose.

FACTORY & OFFICE

Plot No. D 76 , Five Star MIDC Area, Kagal.
Dist. Kolhapur -416216 (M.S.)India.
Email : oxalisptltd@outlook.com

Telefax : 0231-2305072
Phone : 0231-2305062
Mobile : +91 8805867810



Remarks

- ✓ Tryptic Soya Broth(SCDB) with 0.05% SPS blood culture media are for laboratory and professional use only.
- ✓ Not for medicinal use.
- ✓ Do not use bottles that have cracks or defects.
- ✓ Inoculated bottles should be decontaminated prior to discarding. Clinical samples and microbial cultures should be considered as pathogenic biohazard and handled accordingly.
- ✓ Blood drawn for culture must not be allowed to clot. If bacteria are entrapped within a clot, their presence may go undetected.

Disposal

Disposal of infectious material and material that comes in to contact with clinical sample must be decontaminated and dispose of by autoclaving or incineration or established laboratory procedures. User must be ensure safe disposal of used or unusable preparation of the products.

Reference

1. Atlas R. M., 1993, Handbook of Microbiological Media, 147-153, CRC Press, Boca Raton, FL.
2. MacFaddin J. F., 1985, Media for Isolation-Cultivation-Identification-Maintenance of Medical Bacteria, Vol. I, Williams and Wilkins, Baltimore.
3. Murray P. R., Baron J. H., Pfaller M. A., Jorgensen J. H. and Tenover F. C., (Eds.), 2003, Manual of Clinical Microbiology, 8th Ed., American Society for Microbiology, Washington, D.C.
5. Diagnostic Microbiology , Baily & Scott, 9th Edition , Ellen jo Baron, etal. , Mosby 1994.

FACTORY & OFFICE

Plot No. D 76 , Five Star MIDC Area, Kagal.
Dist. Kolhapur -416216 (M.S.)India.
Email : oxalispvtltd@outlook.com

Telefax : 0231-2305072
Phone : 0231-2305062
Mobile : +91 8805867810